



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/666,838	09/17/2003	Stephen C. Beal	022010-000210US	5785

20350 7590 08/08/2006

TOWNSEND AND TOWNSEND AND CREW, LLP
TWO EMBARCADERO CENTER
EIGHTH FLOOR
SAN FRANCISCO, CA 94111-3834

EXAMINER

BROWN, MICHAEL A

ART UNIT	PAPER NUMBER
----------	--------------

3764

DATE MAILED: 08/08/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

88

Office Action Summary	Application No.		Applicant(s)	
	10/666,838		BEAL ET AL.	
	Examiner		Art Unit	
	Michael Brown		3764	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☐ Responsive to communication(s) filed on ____.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-35 is/are pending in the application.
- 4a) Of the above claim(s) ____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) ____ is/are allowed.
- 6) ☒ Claim(s) 1-35 is/are rejected.
- 7) ☐ Claim(s) ____ is/are objected to.
- 8) ☐ Claim(s) ____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on ____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
 Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
 Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. ____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. ____. |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date <u>9-20-04</u> . | 6) <input type="checkbox"/> Other: ____. |

DETAILED ACTION

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1-2, 8-9, 16-17, 20, 23-24, 28-31 and 33-35 are rejected under 35 U.S.C. 102(b) as being anticipated by Nordvik.

Nordvik discloses in figure 2 a system for expressing milk from a human breast, comprising a control system (the entire device) adapted to be worn by a user, a massaging means (8 and an oscillating pump), an expression means 4, containment means (col. 5, lines 41-44), the expression means and the massaging means are operated independently by a control system (valves can control the massage means separate from the expression means or different pumps can operate both independently), the expression means are pneumatically operated (by a pump), pivotable (the member 4 can be pivoted about a secured point of the breast), the containment means includes a pre-sterile disposable container (a baby bottle), the system includes components (values) that are adjustable and the expression and massage patterns may be saved to memory in order to repeat or download to a medical documentation, the system is adapted to express milk from one breast at a time, from both breast at one time or alternatively from each breast (moving the device from one breast to the other or putting one on each breast), the system operates independently or

in combination with valves, pumps and bladders, a method of expressing milk and a rechargeable battery (col. 7, lines 26-30).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 2-7 and 10-15 are rejected under 35 U.S.C. 103(a) as being unpatentable over Nordvik in view of LaTorre, along with Nordvik (fig. 3).

Nordvik discloses in figure 2 a system for expressing milk from a human breast, substantially as claimed, that further comprises a massaging means that includes a series of inflatable rings (9, 12) and a plurality of bladders (6, 10), the bladders are operated pneumatically, arranged in an annular manner (fig. 2) and in a liner manner (they form a curved line) and the expression means is pneumatically operated (by a pump). However, Nordvik doesn't disclose a Velcro fastener used to close a plurality of bladders having an open section. LaTorre teaches in figure 6 a breast cover comprising a Velcro type fastener. Nordvik teaches in figure 3 a system for expressing milk from a human breast comprising an expression means (13, 14), in the shape of a bellows. It would have been obvious to one having ordinary skill in the art at the time that the invention was made that the Velcro fastener as taught by LaTorre could be used to fasten the massage means and the expression means disclosed by Nordvik around the user's breast. The Velcro type fastener would allow the user to open and fasten the

device around the breast. The bellows shaped expression means (fig. 3, which is a different embodiment, different invention) could be substituted for the expression means (fig. 2, different embodiment, different invention) disclosed by Nordvik because both expression means are functionally equivalent.

Claims 18-19, 21-22 and 25-27 are rejected under 35 U.S.C. 103(a) as being unpatentable over Nordvik in view of Palmer.

Nordvik discloses in figure 2 a system for expressing milk from a human breast, substantially as claimed. However, Nordvik doesn't disclose the system being operated automatically, the system being incorporated into a wearable garment, or a controller worn around the user's waist. Palmer teaches in figures 1-3E a system for expressing milk from a human breast comprising a wearable garment 2 and a controller 3 that can be worn around the user's waist. It would have been obvious to one having ordinary skill in the art at the time that the invention was made that the garment and the controller as taught by Palmer could be incorporated into the system disclosed by Nordvik. The garment could be used to hold the device in place and the controller could be used to control the massaging and expression means. The system could operate automatically by the microprocessor as taught by Palmer.

Claim 32 is rejected under 35 U.S.C. 103(a) as being unpatentable over Nordvik in view of Silver '580.

Nordvik discloses in figure 2 a system for expressing milk from a human breast, substantially as claimed. However, Nordvik doesn't disclose a heating means. Silver teaches in figure 3 a system for expressing milk from a human breast comprising a

heating means 10. It would have been obvious to one having ordinary skill in the art at the time that the invention was made that the heating means as taught by Silver could be incorporated into the system disclosed by Nordvik in order to use the heating means to heat the user's breast.

Conclusion

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. No additional prior art was cited in this office action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Michael Brown whose telephone number is 571-272-4972. The examiner can normally be reached on 5:30 am-4:00 pm Monday-Thursday.


If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Gergory Huson can be reached on 571-272-4887. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Application/Control Number: 10/666,838
Art Unit: 3764

Page 6

M. Brown
July 31, 2006

A handwritten signature in black ink, reading "Michael A. Brown". The signature is written in a cursive, flowing style with a long horizontal stroke at the end.

**MICHAEL A. BROWN
PRIMARY EXAMINER**